Update on ISO 10993



Past and Future Changes

What Has Changed?

ISO 10993-1:2018

- Released August 2018
- Significant Changes to Table A.1
 - Moved to informative Annex
 - Six new columns
 - Physical and/or chemical information
 - Material mediated pyrogenicity
 - Chronic toxicity
 - Carcinogenicity
 - Reproductive/developmental toxicity
 - Degradation



Meet the New Standard: Table A.1 from ISO 10993-1

Medical device categorization by		Endpoints of biological evaluation															
Nature of body contact		Contact duration															
Category	Contact	A - limited (≤24 h) B - prolonged (>24 h to 30 d) C - Long term (>30 d)	Physical and/or chemical informa- tion	Cyto toxi city		Irrita tion or intra cuta neous reac tivity	Ma- terial media ted pyro geni city ^a	Acute syste mic toxi city ^b	Sub acu te toxi city ^b	Sub chro nic toxi city ^b	Chr onic toxi city ^b	Impla nta tion ef- fects- b,c	Hem oco mpa tibil ity	Gen otox ici- ty ^d	Car cin oge nic ity ^d	Repro duc- tive/ develop mental toxici- ty ^{d,e}	Deg rada tion ^f
		A	Xs	Eh	E	E											
	Intact skin	В	х	E	E	E											
		С	х	E	Е	E											
Surface medical		A	х	E	E	E											
device	Mucosal membrane	В	х	E	E	E		E	Е			Е					
		С	х	E	Е	E		E	Е	Е	E	E		E			
	Breached or	A	х	E	E	E	E	E									
	compromised	В	Х	E	E	E	E	E	E			E					
	surface	С	х	E	E	E	E	E	Е	E	E	E		E	E		
	Blood path, indirect	A	х	E	E	E	E	E					E				
		В	Х	E	E	E	E	E	E				E				
		С	х	E	E	E	E	E	Е	E	E	E	E	E	E		
Externally	Tissue/	A	Х	E	E	E	E	E									
communicating	bone/	В	х	E	Е	E	E	E	Е			E		Е			
medical device	dentini	С	х	E	E	E	E	E	E	E	E	E		E	E		
		A	х	Е	Е	E	Е	E					E	EĴ			
Circulating blood	Circulating blood	В	х	E	Е	E	E	E	Е			E	E	E			
	С	х	Е	Е	E	E	E	Е	E	E	Е	Е	E	E			

Table A.1 — Endpoints to be addressed in a biological risk assessment









CONFIDENTIAL | © 2019 Sterigenics U.S., LLC | All Rights Reserved.

What Has Changed?

Annex B cannibalized ISO TR 15499:2016

TR 15499:2016 has been "withdrawn". Now proper risk evaluation details are contained within the standard. Transitory contact, nanomaterials, and absorbable materials added

ISO 18562 referenced



ISO 18562

Biocompatibility evaluation of breathing gas pathways in healthcare applications





Why the Changes

Place emphasis on a "Risk Based Approach' instead of checkbox approach

• Emphasis on "Competent Individuals"

To create a thoughtful and planned approach

- Biological Evaluation Plan (BEP)
- Biological Evaluation Report (BER)



Where are With Regulatory Bodies Acceptance?





Where are With Regulatory Bodies Acceptance?





Update on ISO 10993

The Future?



Cytotoxicity 10993-5

MEM vs. MTT/XTT vs. Neutral Red vs. Colony Forming Test

- Round robin being developed to understand differences between tests and labs
- More guidance on failures and investigation approaches



Irritation and Sensitization 10993-10 (23)

Separation of irritation and sensitization

 Sensitization stays as part 10 but irritation becomes part 23

Main irritaiton method is moved to in vitro

• Using Reconstructed Human Epidermis (RHE) tissues



In Vivo Alternatives: Irritation



In Vitro Irritation







Chemical Characterization 10993-18



150/TC 194/SC N Date: 2017-03-09 ISO/CD 10993-18:2017(E) ISO (TO 194 (SC OVE 14 Secretariati DIN Biological evaluation of medical devices - Part 18: Chemical

acterization of materials

et an 190 International Standard. It is distributed for revis lifecut notice and may not be referred to as an Internatio

of this draft are invited to submit, with their comments, notification of any re % they are aware and to provide supporting documentation.

DIS 10993-18 (2019): 79 pages

- Major revision of the whole concept of chemical characterization
- Broader definition of chemical characterization
- •Introduction of the concept of Analytical Evaluation Threshold (AET)
- Definition of extractable testing
- Definition of leachable testing
- Clarification on the extraction procedures and analytical techniques to be used during chemical testing



Chemical Characterization 10993-18

• The whole procedure is described on at the beginning of FDIS ISO 10993-18 (2019)







Applying These Changes

CONFIDENTIAL | © 2018 Nelson Laboratories

Standards for Presentation





CONFIDENTIAL | © 2018 Nelson Laboratories

Past Approach

A Sotera Health company

510(k) Memorandum - #G95-1 Table 1

Initial Evaluation Tests for Consideration

			(Device Categorie		_	0	iote	on te	al	Effe	et	_
			Body Contact (see 4.1)		Contact duration (see 4.2) A-limited (24h) R-prolonged (24h to 30 days) C-permanent (=30days)	Cyntraidy Seathofia Santhofia Libfer e brautoeuu Jacfoly			system (alles) (acted) Self-berg resch (strate resch)		Contracty	ingledica.	
Device	Contact	Perform	Surface devices	Skin Mueos membr Breach compre surface			_,,			ZOLZOLL'			
contact	time	tests	External communicating devices	Blood indirec Tissue/ dentin communicating+	ë	ات ا		1	1	0 2 . 0 0		××1°°°	XXXX
				Circulating blood	A B C	XXX	XXX	XXX	XXX	0 X	o^ N X		XXX
			Tissue/	B C	XXX	XXX	x 0 0	0 0	00	XX		F	
			devices	Blood		XXX	XXX	XXX	xxx	- 0 x	xx	XXX	X X X
Welson	Labs												

Past Approach

510(k) Memorandum - #G95-1 Table 1

Initial Evaluation Tests for Consideration







Medical	device cate	egorization by	61 (1)			
Nature Cor	of Body tact	Body Contact ct Duration				
Category	Contact	A-limited (≤24h) B-prolonged (>24hto30d)	Physical and/or chemical informa- tion	Cytotox- icity	Sensiti- zation	
		(> 30 d)				
	65	A	Xs	Eh	E	
	Intact	В	Х	Е	Е	
	548505.38	С	х	Е	E	
	Mucosal	A	x	Е	E	
Surface	mem-	В	х	E	E	
device	brane	C	Х	E	E	
	Breac	1000		E	Е	
	or com-	В	х	E	Е	
	surface	C	x	Е	E	
	Blood	Α	х	Е	Е	
	path,	В	х	Е	Е	
	indirect	С	х	Е	E	
External	Tissue/	A	X	Е	E	
com mu- nicating	bone/	В	х	Е	E	
device	Gentine	C	Х	E	E	
1		٨	х	Е	Е	
	Circulat- ing blood	В	х	E	E	
	10021419015	C	Х	Е	E	



ISO 10993 and RISK

ISO 10993 is intended as a guidance to determine the potential biological risks arising from the use of medical devices.



Meaning, what is the risk of my materials and processes to the patient?

ISO 10993-1: Biological evaluation of medical devices – Part 1: Evaluation and testing within a **risk** management process



CONFIDENTIAL | © 2018 Nelson Laboratories



Use of International Standard ISO 10993-1, "Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process"

Section III. Risk Management for Biocompatibility Evaluations

"Such a process should generally begin with assessment of the device, including the **material components**, the **manufacturing processes**, the **clinical use of the device**..." Considering this information, the **potential risks from a biocompatibility perspective** should be identified. Considering the potential biological impact, a plan should be developed ... **either by biocompatibility testing or other evaluations that appropriately address the risks**.



Incorporating Risk



What is **Risk**?

ISO 14971 Definition: Combination of the **probability of occurrence** of harm and the **severity of that harm.**



CONFIDENTIAL | © 2018 Nelson Laboratories

Biological Safety Evaluation





CONFIDENTIAL | © 2018 Nelson Laboratories

Biological Evaluation Report















A Sotera Health company

- Biological Risk Assessment
 - Material change
 - Failed test
 - Location change
 - Process change
 - Specific requests by a regulatory

body





Biological Risk Assessment
Material change

BIOLOGICAL RISK ASSESSMENT

Labs

son

CONCLUSION: Based on the literature review of compound/material, it is my opinion that the change would pose a low risk of toxicity and adverse effects to the patient from this material change would be unlikely. Additional animal testing would not generate useful data and would not follow the guidance in ISO 10993 part 2.



- Gap Analysis
- Purpose
 - The purpose of this report is to perform a gap analysis between the completed testing on the device and the current testing requirements. This gap analysis will uncover any testing that may need to be performed to meet the current standards.





- Gap Analysis
- Purpose
 - The purpose of this report is to perform a



GAP ASSESSMENT

CONCLUSION: Based on review of the tests performed and the current standards the biocompatibility of the device is well supported.





More Information





THANK YOU!





CONFIDENTIAL | © 2018 Nelson Laboratories