



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

Nelson Laboratories, LLC

**6280 S. Redwood Road
Salt Lake City, UT 84123**

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

TESTING

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AT-1382

Certificate Number


ANAB Approval

Certificate Valid Through: 03/16/2021
Version No. 009 Issued: 03/12/2020



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Nelson Laboratories, LLC

6280 S. Redwood Road
Salt Lake City, UT 84123

Douglas Fairbanks dfairbanks@nelsonlabs.com
Robert Thoreson rthoreson@nelsonlabs.com
www.nelsonlabs.com 801-290-7500

TESTING

Valid to: **March 16, 2021**

Certificate Number: **AT-1382**

Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Agar Overlay	STP0031 based on ANSI/AAMI/ISO 10993-1,5,12 USP <87>, USP<1031>	Medical Devices, Raw Materials	ISO Class 5 Hoods Microscope Incubators
Antimicrobial Preservative Effectiveness	STP0131 based on USP <51>, STP0132 based on USP <51> and EP 5.1.3	Antimicrobial Preservatives	Incubators
Bacterial Endotoxins	STP0046 based on USP <85>, USP<161>, USP<797>, AAMI ST72, EP 2.6.14, ASTM D7102-04, BS EN 455-3	Medical Devices, Drugs	Microplate Reader
Bacterial Filtration Efficiency (BFE) Viral Filtration Efficiency (VFE)	STP0004 and STP0007 based on ASTM F2101, EN14683, ASTM F2100	Medical & Surgical Face Masks	Andersen Sampler
Viral Penetration and Whole Glove Viral Barrier Testing	STP0062, STP0174, and STP00198 based on ASTM F1671, AAMI PB70, ISO16604, and NFPA 1999	Textiles, Gloves	ISO Class 5 Hoods Incubators
Bioburden	STP0036 based on ISO 11737-1	Textiles, Medical Devices, Tissues, Pharmaceuticals	ISO Class 5 Hoods Incubators

Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Radiation Sterilization Validations and Dose Audits	STP0050 based on ISO 11737-2, 11137-01 and -02, AAMI TIR 17, 35, 37. STP0051 based on ISO 11737-01 and -02, 11137-01 and -02, AAMI TIR 17, 33, 37. STP0195 based on ISO 11137-2 and AAMI TIR 40. STP0044 based on ISO11137-01 and -02, AAMI TIR 33, 35	Textiles, Medical Devices, Tissues, Pharmaceuticals	ISO Class 5 Hoods Incubators
Biological Indicators (Population verification, Process Challenge Device (PCD) Preparation, BI Sterility)	STP0045, STP0079, and SOP0180 based on USP<55>, ISO 11138-1 to -4, ISO 11135-1 to -2, ISO 11138-7 ISO 14937, ISO 17665-2, AAMI TIR 13, 14, 16, BS EN 550	BIs, PCDs	BI Sterility Suite ISO Class 5 Hoods Incubator
Cleaning, Disinfection, Sterilization Including the following sub-analyses (separately accredited): - Hemoglobin - Protein - Carbohydrates - MEM elution - TOC - Bioburden	Template 122, STP0129, STP0194 and Template 202 based on AAMI TIR 12, 30, ASTM E1837, ISO17664, ISO 15883 STP0086 and STP0202 based on ANSI/AAMI ST79, AAMITIR12, ANSI/AAMI/ISO 17665, USP <1211> STP0152 based on AAMI TIR 12, USP<1211>, ANSI/AAMI/ ISO 11135-1 STP0159, Template 124, and Template 194 based on ISO 17664, ANSI/AAMI ST79, ANSI/AAMI ST77, ANSI/AAMI/ISO 11135, AAMI TIR30	Medical Devices, Reusable Devices	Washer/Disinfectors Sterilizers (Steam, EO, VHP) UV/VIS Spectrophotometer
Container Closure Integrity (Bacterial Ingress)	STP0164 based on PDA TR 27 and FDA Guidance for Industry: Container and Closure Integrity Testing	Packaging Materials for Medical Device & Pharmaceutical	Pressure/Vacuum Vessel Incubators

Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Hemolysis	STP0093 based on ANSI/AAMI/ISO 10993-1,4,12 and ASTM F756-08	Medical Devices, Raw Materials	Spectrophotometer Incubators
MEM Elution	STP0032 based on ANSI/AAMI/ ISO 10993-1,5,12 USP <87>, USP<1031>	Medical Devices, Raw Materials	ISO Class 5 Hoods Microscope Incubators
Bacterial Reverse Mutation Assay (Ames Test)	STP0097 and STP0098 based on ISO 10993-1,3,12,33 OECD 471	Medical Devices, Raw Materials	Incubators, Automated Plate Counter
Chromosome Aberration Assay	STP0101 and STP0102 based on ISO 10993-1,3,12,33 OECD 473	Medical Devices, Raw Materials	ISO Class 5 Hoods, Microscope, Incubators
MTT Quantitative Cytotoxicity Test	STP0207 based on ISO10993-5 and ISO10993-12	Medical Devices	Incubator, Microscope, Spectrophotometer
Complement Activation	STP0092 based on ISO 10993-1,4,12	Medical Devices	Spectrophotometer
Partial Thromboplastin Time Test - PTT	STP0094 based on ISO 10993-4, 12 and ASTM F2382	Medical Devices	Incubator
Microbial Retention (Including Filter Bubble Point/Integrity Test)	STP0103 based on ASTM F838-15	Filters	Flow Meter Pressure Gauge ISO Class 5 Hood Incubators
Microbiological Examination of NonSterile Products (Enumeration and Specified Organisms, USP 61/62)	STP0169 and STP0165 based on USP<61> and USP<62>	Medical Devices, Pharmaceuticals	ISO Class 5 Hoods Incubators
Organism Identification (Genetic and Gram Stain)	STP0105, and STP0173 based on USP<1113>	Medical Devices, Pharmaceuticals	Genetic Sequencers Thermocyclers Automatic Gram Stainer ISO Class 5 Hoods Incubators Microscopes
Product Sterility (Cleanroom and Isolator), MPN Method Suitability (Bacteriostasis /Fungistasis), and Isolator Package Validation	STP0077, STP0081, STP0082 and STP0078 based on USP<71>, USP<161>, USP<797>, ISO 11737-2, 11137-01 and -02, PIC/S PI 012-3, EP 2.6.1, JP XV 4.06, ISO 17665, AAMI TIR 33	Medical Devices, Pharmaceuticals, Biologics, Tissues	ISO Class 5 Cleanrooms and Hoods Incubators Isolator

Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Standard Plate Counts	STP0035 based on USP <71> STP0169 based on USP<61>	Water, Food, Cosmetics, Pharmaceuticals	ISO Class 5 Hoods Incubators

Chemical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Ethylene Oxide (EO) Residual Analysis	STP0016 based on ANSI/AMMI/ISO10993-7, 2008, USP <621>	Medical Devices	GC
FTIR, Material Characterization	STP0021 based on USP<851>and USP<197>	Polymers, Non-volatile Residue, Materials	FTIR, Microscope
Water Purity Analysis • TOC • Conductivity • pH	STP0024 and STP0099 based on USP<1231>, USP<1230> and all USP monograph waters, STP0028 based on USP<643> STP0029 based on USP<791> STP0147 based on USP<645>	Water – USP, Water – EP	TOC Analyzer, Conductivity Meter, pH Meter
Biological Marker Analysis • Hemoglobin • Protein • Carbohydrates	STP0087, STP0088, and STP0183 based on ASTM F756-13, AAMI TIR30, and Cleaning, Disinfection, Sterilization references previously listed.	Medical Devices, Reusable Devices	Spectrophotometer
Metals Analysis via Inductively Coupled Plasma – Mass Spectrometry	STP0190 based on USP<233>, and EPA Method 200.8	Medical Devices	Inductively Coupled Plasma – Mass Spectrometer (ICP-MS)
Particulates Testing and VOC Sampling	STP0104 based on ISO 18562-2 and ISO 18562-3	Breathing systems, intubation tubing, other gas pathway devices	DustTrak, Flow meters

Mechanical / Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Barrier Testing: Synthetic Blood and Water Resistance (Hydrostatic Pressure, Impact Penetration)	STP0061, STP0071 and STP0072 based on ASTM F1670, AAMI PB70, ISO 16603, AATCC 42 and 127	Textiles, Gloves	Hydrostatic Head Tester, Incubators
Synthetic Blood Resistance	STP0012 based on ASTM F1862 and ISO 22609	Medical facemasks and surgical respirators	Blood testing apparatus
Flammability	STP0073 based on 16 CFR Part 1610	Face masks, surgical gowns, and surgical drapes	Flammability tester
Container Closure Integrity (Dye Ingress)	STP0149 based on ANSI/AAMI/ISO 11607-1,2, ASTM D4491-07, PDA TR 27 and FDA Guidance for Industry: Container and Closure Integrity Testing	Packaging Materials for Medical Device & Pharmaceutical	Vacuum Vessel, Spectrophotometer
Container Closure Integrity (Mass Extraction)	STP0140 based on ASTM F3287-17	Nonporous rigid containers	ME2 Mass Extraction Leak Test Instrument, Calibrated Leak Orifices
Particulates	STP0011 based on USP <788>, <789>, EP 2.9.19, 2.9.31, BP Appendix XIII A, BH EN 45502-1, 45502-2-1 ISO 8536-4	Medical Devices, Injectables and Ophthalmic Solutions, Pharmaceutical Products	Liquid Particle Counting System, Microscope
Particulate Filtration Efficiency (PFE)	STP0005 based on ASTM F2299	Medical & Surgical Face Masks	Particle Counter, Particle Generator
Respirator Certification Testing (NIOSH N95/N99) - Respirator Inhalation/Exhalation - Respirator Valve Leak - Sodium Chloride Aerosol Test	STP0145 based on 42 CFR Part 84 and NIOSH TEB – APR-STP-007, RCT- APR-STP-003 STP0143 based on 42 CFR Part 84 and NIOSH TEB- APR-STP-0004 STP0014 based on 42 CFR Part 84 and NIOSH TEB- ARP-STP-0058, TEB-ARP-STP-0059	Respirators	Differential Pressure Apparatus, Air Flow Apparatus, Automated Filter Tester, Sodium Chloride Tester, Valve Leak Tester

Mechanical / Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
EN 13795: Performance requirements for surgical gowns and drapes • Microbial penetration resistance (wet and dry) • Microbial evaluation (bioburden) • Particle evaluation • Liquid penetration resistance • Burst strength • Tensile Strength	STP0191 and STP0188 based on EN ISO22610 and EN ISO 22612 STP0036 based on ISO 11737-1 (Bioburden method) STP0144 based on EN ISO 9073-10 STP0071 based on AATCC 127 and EN 20811 STP0192 based on EN ISO 13938-1 STP0066 EN 29073-3	Medical & Surgical Gowns and Drapes	ISO Class 5 Hoods Incubators Gelbo Flex Unit Particle counter Burst tester Instron (Tensile) tester Rulla II testers

Note:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AT-1382.



Vice President